

BLAINE'S AIR PERMEABILITY APPARATUS MODEL NO – SL-CC-062, MAKE – LABTEK, ORIGIN – INDIA



IS 4031, 5516, 1727 and 4828, ASTM C-204 BS 4359-2

The apparatus is used for determining the fineness of cement in terms of specific surface expressed as total surface area in square centimeters per gram of cement. This is a variable flow type air permeability.