

COMPACTION FACTOR APPARATUS

MODEL NO SL -CC -075, MAKE - LABTEK, ORIGIN - INDIA



IS 1199, 5515. & BS 1881-103

The apparatus is used for determining the work ability of fresh concrete, provided the maximum size of the aggregate does not exceed 38mm. The test is particularly useful for concrete mixes of very low work ability where true slump values are not reliable.