

CURING TANK

MODEL NO – SL-CC-111, MAKE – LABTEK, ORIGIN – INDIA



- a) 24 Hour cycle from time of mixing.
- b) Controlled 35oC or 100oC \pm 2oC Curing Temperature for concrete.
- c) Controlled 27oC \pm 2oC Curing Temperature for grey cement.