

VOLUME CHANGE APPARATUS

MODEL NO – SL-CC-087, MAKE – LABTEK, ORIGIN – INDIA



ASTM C 490, IS 4031 & BS 1881

The instrument is used for determining the volume change of cement concrete.

Specification :

The apparatus comprises of one mould effective gauge length complete with base plate, four reference pins, one length comparator frame, one stainless reference bar with insulated grip, and one dial gauge, 0.002mm x 10mm